

EU DECLARATION OF CONFORMITY

Manufacturer: **Mercator Medical (Thailand) Ltd.**
88/8 Moo.12 Tambon Kampaengphet,
Amphur Rattaphum, Songkhla 90180 Thailand

Declare under its sole responsibility that protective gloves:

| Brand | Type | Sizes | Reference Numbers |
|----------------------|---|-----------------------|---------------------|
| MERCATOR gogrip long | Nitrile, powder free, blue, diamond textured, long cuff, for single use | S (6-7) - XXL (10-11) | RP300310 02-06_0001 |

The products described above are in conformity with the provision of the PPE Regulation (EU) 2016/425 as a Category III product, Type B where such is the case, with standards EN ISO 374-1:2016+A1:2018, EN ISO 374-2:2019, EN ISO 374-4:2019, EN ISO 374-5:2016, EN 16523-1:2015+A1:2018 and EN ISO 21420: 2020

The products are subject of EU Type Examination (Module B) and Module C2 on-going Conformity under certificate No. 2777/14031-07/E00-00 issued by notified body:

SATRA Technology Europe Limited (Notified Body number 2777)
Bracetown Business Park, Clonee Dublin 15, Dublin, Ireland

This declaration is supported by the Quality System approval to ISO13485:2016/EN ISO13485:2016 and ISO 9001:2015 issued by:

SGS United Kingdom Ltd
Rossmore Business Park Ellesmere Port Cheshire CH65 3EN UK

Date and place of issue:
02.02.2023, Thailand

Signed on the behalf of the Manufacturer.

Mr. Praneet Inthajak

Senior QA/RA Manager

Mercator Medical (Thailand) Ltd.

**Mercator Medical (Thailand) LTD.**

88/8 Moo 12
Tambon Kampaengphet
Amphur Rattaphum
Songkhla 90180 Thailand

ph.: +66 74 584 222
fax.: +66 74 584 223
enquiry@mercator.co.th
www.mercator.co.th

Sales Office
Mercator Medical S.A.
ul. H. Modrzewskiej 30
31-327 Kraków, Poland

ph.: +48 (12) 66 55 480
fax.: +48 (12) 66 55 415
medical@mercatormedical.eu
www.mercatormedical.eu